KAVALIER

CERTIFICATE OF CONFORMITY

No. KAG/92/2020

Issuer's name: **KAVALIERGLASS, a.s.**

Issuer's address: Křížová 1018/6, Praha 5
office: Sklářská 359, 285 06 Sázava, Czech Republic

Object of the declaration: **LABORATORY GLASSWARE**

- **CYLINDERS GRADUATED 1634 / AM**



Material: Borosilicate glass **SIMAX®**

Purpose of use: laboratory volumetric glassware

The object of the declaration described above is in conformity with the requirements of the following documents:

- **ISO 3585 Borosilicate glass 3.3 - Properties.**
- **ISO 4788 Laboratory glassware - Graduated measuring cylinders.**

- Maximum permissible errors in capacity fulfill the values specified in Table 1/ISO 4788.

These errors represent the maximum permissible error at any point on the scale.

Table 1 - ISO 4788:

Nominal capacity ml		**5**	**10**	**25**	**50**	**100**	**250**	**500**	**1000**	**2000**
Max. permissible error ± ml	**Class A**	0,05	0,1	0,25	0,5	0,5	1,0	2,5	5,0	10,0

The producer confirms hereby that the characteristics, measures and accuracy of the products listed above are in full conformity with the provisions of the standard.

Additional information:

The producer also declares that the products are safe when used in usual and proper way.

The producer has installed the Quality Assurance System according to ISO 9001 and thus guarantees that all products delivered to the market are in full conformity with the technical documentation and with all fundamental requirements to such products.

Certificate No. 04100940602 issued by TÜV CERT, Certification Body at TÜV NORD CERT GmbH.

Sázava, 24. 04. 2020

Place and date of issue

Ing. Viera Košíková

quality manager

KAVALIERGLASS, a.s.
Křížová 1018/6, 150 00 Praha 5
office: Sklářská 359, 285 06 Sázava
Czech Republic
IČ: 474 68 815